



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/597,893	05/20/2008	Robert Graham Urie	EPCLP0122US	7532

23908 7590 06/07/2010  
RENNER OTTO BOISSELLE & SKLAR, LLP  
1621 EUCLID AVENUE  
NINETEENTH FLOOR  
CLEVELAND, OH 44115

EXAMINER
----------

LLOYD, EMILY M

ART UNIT	PAPER NUMBER
----------	--------------

3736

MAIL DATE	DELIVERY MODE
-----------	---------------

06/07/2010

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/597,893	<b>Applicant(s)</b> URIE, ROBERT GRAHAM	
	<b>Examiner</b> EMILY M. LLOYD	<b>Art Unit</b> 3736	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 10 August 2006.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-25 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☐ Claim(s) \_\_\_\_\_ is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☒ Claim(s) 1-25 are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |   |   |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                    | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)         | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____   | 6) <input type="checkbox"/> Other: _____                          |

Art Unit: 3736

## DETAILED ACTION

### *Election/Restrictions*

1. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 1-10, 18, 20 and 21, drawn to a guidewire (and kit containing a guidewire) where stiffness is defined as the force required to produce an angular lateral displacement of 30 degrees.

Group II, claim(s) 11-17, 19 and 22-24, drawn to a guidewire (and a kit containing a guidewire) where stiffness is interpreted as it is usually used in the art (and without the special definition of Group I).

Group III, claim(s) 25, drawn to a method using a needle, a guidewire, and a peelable catheter sheath, where the peelable catheter sheath has a distal end with an outer diameter of less than approximately 2 mm and a proximal end having an outside diameter of at least 4 mm.

2. The groups of inventions listed above do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

Group I requires a guidewire with specific stiffnesses, where stiffness is defined by Applicant as "the force required to produce an angular lateral displacement of 30 degrees"; Groups II and III do not use this definition of stiffness. Group II requires a guidewire with portions having different stiffnesses, where stiffness is defined as it is usually used in the art (and without the special definition of Group I); Group I uses a different definition of "stiffness" and Group III does not set forth any limitations with regards to stiffness. Group III requires details regarding the peelable catheter sheath that are not in Groups I and II, and Group III also does not set forth any stiffness limitations with regards to stiffness as claimed in either of Groups I or II.

Art Unit: 3736

3. The Examiner further notes that it appears that claim 18 may be intended to depend on claim 17 and claim 20 may be intended to depend on claim 11. If amended to change the dependency, claims 18 and 20 would be part of Group II instead of Group I.

4. A telephone call was made to Don Bulson on 1 June 2010 to request an oral election to the above restriction requirement, but did not result in an election being made.

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement may be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse. Traversal must be presented at the time of election in order to be considered timely. Failure to timely traverse the requirement will result in the loss of right to petition under 37 CFR 1.144. If claims are added after the election, applicant must indicate which of these claims are readable on the elected invention or species.

Should applicant traverse on the ground that the inventions have unity of invention (37 CFR 1.475(a)), applicant must provide reasons in support thereof. Applicant may submit evidence or identify such evidence now of record showing the

Art Unit: 3736

inventions to be obvious variants or clearly admit on the record that this is the case.

Where such evidence or admission is provided by applicant, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C. 103(a) of the other invention.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to EMILY M. LLOYD whose telephone number is (571)272-2951. The examiner can normally be reached on Monday through Friday 8:30 AM - 5 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on 571-272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Emily M Lloyd  
Examiner  
Art Unit 3736

Application/Control Number: 10/597,893

Page 5

Art Unit: 3736

/EML/

/Max Hindenburg/

Supervisory Patent Examiner, Art Unit 3736